



Figure 1 Transformer T<sub>1</sub> isolates the temperature sensor, IC<sub>2</sub>, from the equipment under test. The period of IC<sub>1</sub>'s digital output varies as a function of temperature. The circuit's output period varies at a rate of 10 μsec/°K. User-selected scale factors range from 10 to 640 μsec/°K.